

**HAZARD MANAGEMENT SAFE OPERATING PROCEDURE (SOP)**

Only to be completed where required as a control measure under a Risk Assessment

NAME OF THE TASK/ACTIVITY	NIKON SMZ25 STEREO FLUORESCENCE MICROSCOPE	DATE: 12/02/2020
LOCATION	ADELAIDE MICROSCOPY WAITE FACILITY BUILDING 19 ROOM G5 ADELAIDE MICROSCOPY FROME ROAD FACILITY HELEN MAYO NORTH	Insert photo (Optional)
<b>RISK ASSESSMENT (RA) NAME</b>	Stereo Microscope Nikon SMZ25_RA_Waite	

Residual risk rating from the RA                      Low                      Medium                      High                      Very High

Procedural controls:

Only trained users to operate the instrument. All new users are to be given practical training by a member of Adelaide Microscopy staff. Users must also follow guidelines safe the main panel for operation of the microscope and UV light source.

Refer to risk assessment and safe operating procedure for mercury lamps. Do not use the lamp beyond its life span. A warning is used over the life expectancy of the lamp is a risk that the lamp may burst and leak mercury vapour. A

HSW Handbook	Hazard Management	Effective Date:	17 December 2019	Version 3.0
Authorised by	Chief Operating Officer (University Operations)	Review Date:	17 December 2022	Page 2 of 2
Warning	This process is uncontrolled when printed. The current version of this document is available on the HSW Website.			